



### Code 6207 HYDROCHECK SPECTRATEST CHLORINE (DPD)

RANGE: 0,1 - 1,5 ppm Cl<sub>2</sub>

NUMBER OF TEST: 400

#### KIT CONTAINS:

- 2 bottle reagent Chlorine A
- 1 bottle reagent Chlorine B
- 2 bottle reagent Chlorine C
- 2 glass tubes 16 mm with stopper
- 2 test tubes 20cc
- 1 measuring syringe for sample
- instructions

#### NOTES:

The SPECTRATEST kits must be used with a spectrophotometer with a 16-mm-round cell-holder.

It is advisable to use a spectrophotometer as the HC 7800 or Uvi-Line, because we provide them with preloaded factors and they do not require the preparation of a calibration curve.

The SPECTRATEST kits, which offer a much lower cost per analysis, are recommended when performing many analysis to be repeated frequently.

#### REFILLS

refill spectratest Chlorine

Code 6422

- 3 bottles reagent Chlorine A
- 3 bottles reagent Chlorine B
- 3 bottles reagent Chlorine C

test tube 20cc, package 10

Code 3310

glass tube 16 mm, package 24

Code 7132