



Code 6235 HYDROCHECK SPECTRATEST SILICA

RANGE: 0,1 – 2 ppm SiO_2

NUMBER OF TEST: 200

KIT CONTAINS:

- 2 bottles reagent Silica A
- 2 bottles reagent Silica B
- 2 bottles reagent Silica C
- 2 glass tubes 16 mm with stopper
- 2 test tubes 20cc
- 1 measuring syringe for sample
- instructions

NOTES:

The SPECTRATEST kits must be used with a spectrophotometer with a 16-mm-round cell-holder.

It is advisable to use a spectrophotometer as the HC 7800 or Uvi-Line, because we provide them with preloaded factors and they do not require the preparation of a calibration curve.

The SPECTRATEST kits, which offer a much lower cost per analysis, are recommended when performing many analysis to be repeated frequently.

REFILLS

refill spectratest silica

Code 6475

- 3 bottles reagent Silica A
- 3 bottles reagent Silica B
- 3 bottles reagent Silica C

test tube 20cc, package 10

Code 3310

glass tube 16 mm, package 24

Code 7132