



## 6120 HYDROCHECK COLORTEST MOLIBDENUM

**METHOD OF ANALYSIS:** Measurement of the intensity of the yellow color developed by the reaction of Molybdenum previously oxidized with a derivative of Mercaptoacetic acid.

**TYOLOGY:** colorimetric analysis

**MEASURING RANGE:** 5 – 10 – 20 – 40 – 60 – 80 – 100 - 150 mg/l Mo

**NUMBER OF ANALYSIS:**150

**IDEAL FOR WATERS:** purified, wastewater and thermal circuits.

### PACKAGE CONTENTS:

- 1 bottle of 15 g of reagent A
- 1 bottle of 10 ml of reagent B
- 1 measuring spoon n° 3
- 1 syringe of 5 ml
- 2 test tubes of 20 ml with stopper
- 1 optical comparator
- 1 chromatic scale
- instructions (English)

### INTERFERENCE:

The substances normally present in the water do not interfere.

### NOTE:

- The optical comparator contained in the package allows the comparison of the color between two test tubes containing the same amount of water, so as to improve the accuracy of the result.

REFFIL or ACCESSORIES	PACKAGE	CODE
Reagent	3 x 15g reag. A + 3 x 10ml reag. B	6375
Plastic tubes	10 x 20 ml	3310