



Code N° 1023 IPT SWIMMING POOLS (free chlorine)

SENSITIVITY: 0,2 pH unit / 0,1 ppm Cl_2

TEST QUANTITY: 120 / 150

KIT CONTAINS:

- 1 bottle 15cc reagent Chlorine DPD A
- 1 bottle 15cc reagent Chlorine DPD B
- 1 bottle 15cc reagent pH 6,5-8,5
- 1 test tube 10cc with stopper
- 1 syringe 5cc
- 1 color scale
- use directions

NOTES:

- Normally swimming pools contain organic substances which can react with chlorine to form chloramine. Therefore it is advisable to control free chlorine, combined chlorine and total chlorine using the kit 1011 IPT CHLORINE DPD.

SPARE PARTS:

- reagent A, 6 bottles 15cc
- reagent B, 6 bottles 15cc
- reagent pH, 6 bottles 15cc
- test tube 10cc, 12
- 1 color scale Chlorine DPD,
- 1 color scale pH 6,8-8,4

Code N° 3239
Code N° 3240
Code N° 3208
Code N° 3312
Code N° 3405
Code N° 3415