



Code N° 1024 IPT SILICA

SENSITIVITY: 0,25 ppm SiO_2

TEST QUANTITY: 50

KIT CONTAINS:

- 1 bottle 15cc reagent Silica A
- 1 bottle 15cc reagent Silica B
- 1 bottle 15cc reagent Silica C
- 2 test tube 10cc with one stopper
- 1 color scale 0,25-8 ppm
- comparator
- use directions

NOTES:

- The method is very sensitive. Using an extended layer in a glass tube (like Nessler tube) even lower concentrations can be easily recognized
- For higher precision we recommend the HYDROCHECK COLORTTEST (color comparison), HYDROCHECK MONOTEST or SPECTRATEST (both for use with a spectrophotometer).

SPARE PARTS:

- reagent Silica A, 6 bottles 15cc
- reagent Silica B, 6 bottles 15cc
- reagent Silica C, 6 bottles 15cc
- test tube 10cc, 12
- 1 color scale 0,1-1,5 ppm,

Code N° 3218
Code N° 3219
Code N° 3220
Code N° 3312
Code N° 3404