



## Code 6233 HYDROCHECK SPECTRATEST COPPER

RANGE: 0,25 – 5 ppm Cu

NUMBER OF TEST: 100

### KIT CONTAINS:

- 2 bottles reagent Copper A
- 1 bottle reagent Copper B
- 2 glass tubes 16 mm with stopper
- 2 test tubes 20cc
- 1 measuring syringe for sample
- instructions

### NOTES:

The SPECTRATEST kits must be used with a spectrophotometer with a 16-mm-round cell-holder.

It is advisable to use a spectrophotometer as the HC 7800 or Uvi-Line, because we provide them with preloaded factors and they do not require the preparation of a calibration curve.

The SPECTRATEST kits, which offer a much lower cost per analysis, are recommended when performing many analysis to be repeated frequently.

### REFILLS

refill spectratest copper

Code 6470

- 6 bottles reagent Copper A
- 3 bottles reagent Copper B

test tube 20cc, package 10

Code 3310

glass tube 16 mm, package 24

Code 7132