



## Code 6203 HYDROCHECK SPECTRATEST CYANIDE

RANGE: 0,05 - 0,5 ppm CN<sup>-</sup>

NUMBER OF TEST: 60

### KIT CONTAINS:

- 1 bottle reagent Cyanide A (powder)
- 1 bottle reagent Cyanide B (powder)
- 1 bottle reagent Cyanide C
- 1 spoon n° 1
- 1 spoon n° 2
- 2 glass tubes 16 mm with stopper
- 2 test tubes 20cc
- 1 measuring syringe for sample
- instructions

### NOTES:

The SPECTRATEST kits must be used with a spectrophotometer with a 16-mm-round cell-holder.

It is advisable to use a spectrophotometer as the HC 7800 or Uvi-Line, because we provide them with preloaded factors and they do not require the preparation of a calibration curve.

The SPECTRATEST kits, which offer a much lower cost per analysis, are recommended when performing many analysis to be repeated frequently.

### REFILLS

refill spectratest cyanide

Code 6405

- 3 bottles reagent Cyanide A (powder)
- 3 bottles reagent Cyanide B (powder)
- 3 bottles reagent Cyanide C

test tube 20cc, package 10

Code 3310

glass tube 16 mm, package 24

Code 7132