



Code N° 1001 IPT M ACIDITY

SENSITIVITY: 0,5 french degrees (on 10 ml)

TEST QUANTITY: 625 french degrees

KIT CONTAINS:

- 1 bottle 25ml titrant Acidity
- 1 bottle 15ml indicator M
- 1 test tube 20ml
- use directions

NOTES:

The classic indicator (methylorange) has been replaced by our “indicator m”, which has a much better colour change at the end point.

For higher precision we recommend the HYDROCHECK kits.

SPARE PARTS:

- | | |
|------------------------------------|--------------|
| • titrant acidity, 6 bottles 25ml | Code N° 3105 |
| • titrant acidity, 6 bottles 250ml | Code N° 3106 |
| • indicator, 6 bottles 15ml | Code N° 3003 |
| • 10 test tube 20ml, | Code N° 3310 |

INTERFERENCE:

Water colors such as to prevent the appreciation of the color changes of the indicators.

Low water temperatures slow down the color change of the indicator.

The recommended temperature for the analysis is between 20 and 30 ° C.