



Code N° 1012 IPT CHLORIDE (titration with mercury nitrate)

SENSITIVITY: 0,5 french degrees on 10 ml (7,1 mg/L Cl)

TEST QUANTITY: 625 french degrees

KIT CONTAINS:

- 1 bottle 25ml titrant chloride
- 1 bottle 15ml reagent chloride A
- 1 bottle 15ml reagent chloride B
- 1 test tube 20ml
- use directions

NOTES:

As the titration has a threshold of about 2 mg/L, for low chloride concentrations we recommend the colorimetric method (our IPT N. 1031)

For higher precision we recommend the HYDROCHECK kits.

SPARE PARTS:

- | | |
|--|--------------|
| • titrant chloride, 6 bottles 25ml | Code N° 3109 |
| • titrant chloride 1 °f, 6 bottles 250ml | Code N° 3110 |
| • reagent chloride A, 6 bottles 15ml | Code N° 3001 |
| • reagent chloride B, 6 bottles 15ml | Code N° 3203 |
| • 10 test tube 20ml, | Code N° 3310 |

INTERFERENCE:

Water colors such as to prevent the appreciation of the color changes of the indicators.

Low water temperatures slow down the color change of the indicator.

The recommended temperature for the analysis is between 20 and 30 ° C.