Code N° 1019 IPT NITRITE

SENSITIVITY: 0,05 ppm NO_2

RANGE: 0,05 - 0,1 - 0,2 - 0,4 - 0,6 - 0,8 - 1,0 - 1,5 ppm NO_2

TEST QUANTITY: 50

KIT CONTAINS:

- 1 bottle 25ml reagent Nitrite A
- 2 test tubes 10ml with one stopper
- 1 color scale
- comparator
- use directions

NOTES:
For higher precision we recommend the HYDROCHECK COLORTEST (color comparison), HYDROCHECK MONOTEST or SPECTRATEST (both for use with a spectrophotometer).

SPARE PARTS:

- reagent A, 6 bottles — Code N° 3216
- 12 test tube 10ml, — Code N° 3312
- 1 color scale — Code N° 3410

INTERFERENCE
Chromates and Iron > 1 ppm.
Molybdenum and Sulphides > 10 ppm.
Calcium, Manganese, Lead and Copper > 1000 ppm.
The other substances normally present in the water do not interfere.