

IPT COPPER

Directions for use

1. Rinse the measuring vessel with the water to be tested and fill to the 5 ml mark.
2. Add 10 drops reagent Copper A and swirl
3. Add 5 drops reagent Copper B and swirl again
4. After 5 minutes put the test tube standing on the colour scale and find the colour which matches best.